



Biomet Orthopedics
56 East Bell Drive
P.O. Box 587
Warsaw, Indiana 46581 USA

01-50-0907

Revision B

Date: 2014-09



Biomet Reach Revision Hip System

ATTENTION OPERATING SURGEON

BIOMET REACH REVISION HIP SYSTEM IS NOT INDICATED FOR PRIMARY USE WITHIN THE EUROPEAN UNION

DESCRIPTION

The Biomet Reach Revision Hip system includes either a one-piece femoral component or a modular proximal body that can be used with Biomet M-H modular distal components. The system is used with Type I taper modular heads and compatible acetabular shells/liners and screws. Components are intended for non-cemented revision procedures.

MATERIALS

Femoral Stems	CoCrMo Alloy or Titanium Alloy
Femoral Proximal Bodies	Titanium Alloy
Porous Coating	Titanium Alloy
Auxillary Implants	Titanium Alloy or CoCrMo Alloy

INDICATIONS

1. Noninflammatory degenerative joint disease including osteoarthritis and avascular necrosis.
2. Rheumatoid arthritis.
3. Correction of functional deformity.
4. Revision procedures where other treatment or devices have failed.
5. Treatment of non-union, femoral neck fracture, and trochanteric fractures of the proximal femur with head involvement, unmanageable using other techniques.

Porous coated devices are marketed for non-cemented use in the United States for skeletally mature patients undergoing hip replacement surgery as a result of non-inflammatory degenerative joint disease.

CONTRAINDICATIONS

Absolute contraindications include: infection, sepsis, and osteomyelitis.

Relative contraindications include: 1) uncooperative patient or patient with neurologic disorders who are incapable of following directions, 2) osteoporosis, 3) metabolic disorders which may impair bone formation, 4) osteomalacia, 5) distant foci of infections which may spread to the implant site, 6) rapid joint destruction, marked bone loss or bone resorption apparent on roentgenogram, and 7) vascular insufficiency, muscular atrophy, or neuromuscular disease.

WARNINGS

Improper selection, placement, positioning, alignment and fixation of the implant components may result in unusual stress conditions which may lead to subsequent reduction in the service life of the prosthetic components. Malalignment of the components or inaccurate implantation can lead to excessive wear and/or failure of the implant or procedure. Inadequate preclosure cleaning (removal of surgical debris) can lead to excessive wear. Improper preoperative or intraoperative implant handling or damage (scratches, dents, etc.) can lead to crevice corrosion, fretting, fatigue fracture and/or excessive wear. Do not modify implants. The surgeon is to be thoroughly familiar with the surgical technique, implants and instruments, prior to performing surgery.

1. Use Biomet femoral and modular head component with appropriate matching "Type I Taper".
2. Firmly seat modular head components to minimize the risk of dissociation. Thoroughly clean and dry taper prior to attachment of the modular head component to minimize the risk of crevice corrosion and improper seating.
3. Tight fixation of all non-cemented components at the time of surgery is critical to the success of the procedure. Each component must properly press fit into the host bone which necessitates precise operative technique and the use of specified instruments. Bone stock of adequate quality must be present and appraised at the time of surgery.
4. Complete preclosure cleaning and removal of metallic debris and other surgical debris at the implant site is critical to minimize wear of the implant articular surfaces.
5. Patient smoking may result in delayed healing, non-healing and/or compromised stability in or around the placement site.

Biomet joint replacement prostheses provide the surgeon with a means of reducing pain and restoring function for many patients. While these devices are generally successful in attaining these goals, they cannot be expected to withstand the activity levels and loads of normal, healthy bone and joint tissue.

Accepted practices in postoperative care are important. Failure of the patient to follow postoperative care instructions involving rehabilitation can compromise the success of the procedure. The patient is to be advised of the limitation of the reconstruction and the need for protection of the implants from full load bearing until adequate fixation and healing have occurred. Excessive, unusual and/or awkward movement and/or activity, trauma, excessive weight, and obesity have been implicated with premature failure of certain hip implants by loosening, fracture, dislocation, subluxation and/or wear. Loosening of the implants can result in increased production of wear particles, as well as accelerate damage to bone making successful revision surgery more difficult. The patient is to be made aware in advance and warned of general surgical risks, possible adverse effects as listed, and to follow the instructions of the treating physician including follow-up visits.

Patient selection factors to be considered include: 1) need to obtain pain relief and improve function, 2) ability and willingness of the patient to follow instructions, including control of weight and activity level, 3) a good nutritional state of the patient, and 4) the patient must have reached full skeletal maturity.

Do not reuse implants. While an implant may appear undamaged, previous stress may have created imperfections that would reduce the service life of the implant. Do not treat patients with implants that have been, even momentarily, placed in a different patient.

Device is single use only. After use, the device may be a potential biohazard. Reuse of devices labeled for single-use may result in product contamination, patient infection and/or failure of the device to perform as intended.

PRECAUTIONS

Specialized instruments are designed for Biomet joint replacement systems to aid in the accurate implantation of the prosthetic components. The use of instruments or implant components from other systems can result in inaccurate fit, incorrect sizing, excessive wear and device failure. Intraoperative fracture or breaking of instruments has been reported. Surgical instruments are subject to wear with normal usage. Instruments that have experienced extensive use or excessive force, are susceptible to fracture. Surgical instruments should only be used for their intended purpose. Biomet recommends that all instruments be regularly inspected for wear and disfigurement, prior to surgery.

All trial, packaging, and instrument components must be removed prior to closing the surgical site, do not implant.

POSSIBLE ADVERSE EFFECTS

1. Material sensitivity reactions. Implantation of foreign material in tissues can result in histological reactions involving various sizes of macrophages and fibroblasts. The clinical significance of this effect is uncertain, as similar changes may occur as a precursor to or during the healing process. Particulate wear debris and discoloration from metallic and polyethylene components of joint

implants may be present in adjacent tissue or fluid. It has been reported that wear debris may initiate a cellular response resulting in osteolysis or osteolysis may be a result of loosening of the implant. Further, there has been a report dated March 2010, titled "Advice from the CSD Expert Advisory Group on the biological effects of metal wear debris generated from hip implants," regarding an association between articulating surfaces of: 1) CoCrMo alloy on CoCrMo alloy, 2) CoCrMo alloy on polyethylene, and 3) Titanium alloy on polyethylene in hip replacements and increased genotoxicity. This report, however, did not assess either the clinical relevance of the data or make any definite conclusions as to which metal ions or interactions between metal ions or particulate metals might be responsible for the observed data. The report further cautioned that an association does not necessarily mean a causal relationship, and that any potentially increased risk associated with metal ions needs to be balanced against the benefits resulting from hip replacement. Foreign body sensitivity, where suspected, metal sensitivity tests are to be made prior to implantation.

2. Early or late postoperative infection and/or allergic reaction.
3. Intraoperative bone perforation or fracture may occur, particularly in the presence of poor bone stock caused by osteoporosis, bone defects from previous surgery, bone resorption, or while inserting the device.
4. Loosening or migration of the implants can occur due to loss of fixation, trauma, malalignment, malposition, bone resorption, or excessive unusual and/or awkward movement and/or activity.
5. Periarticular calcification or ossification, with or without impediment of joint mobility.
6. Inadequate range of motion due to improper selection or positioning of components.
7. Undesirable shortening of limb.
8. Dislocation and subluxation due to improper selection, inadequate fixation, malalignment, malposition, excessive, unusual and/or awkward movement and/or activity, trauma, weight gain, or obesity. Muscle and fibrous tissue laxity can also contribute to these conditions.
9. Fatigue fracture of component can occur as a result of loss of fixation, strenuous activity, malalignment, trauma, non-union, or excessive weight.
10. Fretting and crevice corrosion can occur at interfaces between components.
11. Wear and/or deformation of articulating surfaces.
12. Trochanteric avulsion or non-union as a result of excess muscular tension, early weight bearing, or inadequate reattachment.
13. Problems of the knee or ankle of the affected limb or contralateral limb aggravated by leg length discrepancy, too much femoral medialization or muscle deficiencies.
14. Postoperative bone fracture and pain.

MRI INFORMATION

The Biomet Reach Revision Hip System has not been evaluated for safety and compatibility in the MR environment. The devices have not been tested for heating or migration in the MR environment. The risks associated with a passive implant in an MR environment have been evaluated and are known to include heating, migration, and image artifacts at or near the implant site.

STERILITY

Prosthetic components are sterilized by exposure to a minimum dose of 25 kGy of gamma radiation. Do not resterilize. Do not use any component from an opened or damaged package. Do not use implants after expiration date. Single Use Only. Do not Reuse.

CAUTION: Federal law (USA) restricts this device to sale, by or on the order of a physician.

Comments regarding this device can be directed to Attn: Regulatory Dept., Biomet, Inc., P.O. Box 587, Warsaw, IN 46581 USA, Fax 574-372-3968.

All trademarks herein are the property of Biomet, Inc. or its subsidiaries unless otherwise indicated.

CE Mark on the package insert (IFU) is not valid unless there is a CE Mark on the product (description) label.



Authorized Representative: Biomet, U.K., Ltd.
Waterton Industrial Estate,
Bridgend
CF31 3XA U.K.



Symbol Legend	
	Manufacturer
	Date of manufacture
	Do not reuse
	Do not resterilize
	Caution, see instructions for use
	Sterilized using ethylene oxide
	Sterilized using irradiation
	Sterile
	Sterilized using aseptic processing techniques
	Sterilized using steam or dry heat
	Rx Only Caution: Federal law (USA) restricts this device to sale by or on the order of a physician.
	Do not use if package is damaged (Pack Damaged)
	Use by date
	WEEE device
	Catalogue number
	Batch code
	Flammable
	Authorized representative in the European Community